



## CERVICAL SPINE TRUSS SYSTEM PLATING SOLUTION (CSTS-PS)

### INSTRUCTIONS FOR USE (ENGLISH)

#### PLEASE READ CAREFULLY

#### CAUTION: FEDERAL LAW (USA) RESTRICTS THESE DEVICES TO SALES BY OR ON THE ORDER OF A PHYSICIAN.

#### INTENDED USE:

The Cervical Spine Truss System Plating Solution (CSTS-PS) is intended to provide supplemental fixation for the Cervical Spine Truss System (CSTS) Interbody Fusion Device in the cervical spine while biologic fusion takes place.

#### DEVICE DESCRIPTION:

The Cervical Spine Truss System Plating Solution (CSTS-PS) is comprised of cervical plates and screws. The cervical plates have a rotating locking tab for each double screw position to prevent back-out of the screw. The plates are available in 1-level, 2-level, 3-level, 4-level, and 5-level configurations. Each plate is available in multiple lengths to accommodate varying patient anatomy. The screws are available in two diameters and various lengths. All CSTS-PS plates and screws are made from Ti6Al4V alloy.

#### INDICATIONS FOR USE:

The Cervical Spine Truss System Plating Solution (CSTS-PS) is intended for anterior interbody screw fixation of the cervical spine at levels C2-T1.

The Cervical Spine Truss System Plating Solution (CSTS-PS) is used in temporary stabilization of the anterior spine during the development of cervical spinal fusions in patients with degenerative disc disease (as defined by neck pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), trauma (including fractures), tumors, deformity (defined as kyphosis, lordosis or scoliosis), pseudoarthrosis and/or failed previous fusions.

WARNING: This device is not approved for screw attachment to the posterior elements (pedicles) of the cervical, thoracic, or lumbar spine.

#### CONTRAINDICATIONS:

- The CSTS-PS should not be implanted in patients with:
  - An active infection at the operative site or other active systemic infections
  - Tumor involvement at the operative site
  - Rapid progressive joint disease, bone absorption, osteopenia, and/or osteoporosis
  - Known sensitivity to the material
  - Any medical or surgical condition which would preclude the potential benefit of surgery

#### WARNINGS AND PRECAUTIONS:

- Safety and effectiveness have not been established in the following conditions:
  - Gross obesity
  - Smoking
  - Symptomatic cardiac disease
  - Pregnancy
  - Significant loss of bone stock as seen with osteoporosis or osteomalacia
  - Conditions requiring chronic corticosteroid use
  - Active drug abuse

The CSTS-PS is for single use only.

Bending or fracture of the implants or instruments can occur if not handled properly.

Excessive torque applied to the screws when seating the plate may strip the threads in the bone.

No implant should be reused if it has come in contact with blood or other bodily fluids.

All instrumentation and certain implants are provided non-sterile and must be steam sterilized prior to use.

All implants and instrumentation should be inspected prior to use for possible damage or defects. Any damaged or defective component should not be used and should be returned to 4WEB.

Interbody fusion devices are intended to provide mechanical support while biologic fusion occurs. In the event of pseudoarthrosis or delayed fusion, the risk of implant migration, loosening or breakage increases. The physician/surgeon should consider the levels of implantation, patient weight, patient activity level, other patient conditions, etc. which may impact the performance of the system.

Only surgeons trained in cervical spine fusion procedures should implant this device.

#### WARNING: THIS DEVICE IS NOT INTENDED FOR SCREW ATTACHMENT OR FIXATION TO THE POSTERIOR ELEMENTS (PEDICLES) OF THE CERVICAL, THORACIC, OR LUMBAR SPINE.

#### OPERATIVE PRECAUTIONS:

The surgeon is to be thoroughly familiar with the CSTS-PS, methods of application, instruments and surgical technique. The CSTS-PS should be implanted singly. Correct positioning of the CSTS-PS relative to the vertebrae should be checked intraoperatively with x-ray. The height of the CSTS-PS must be chosen on the basis of the patient's anatomy and desired correction provided by the interbody fusion device. Implantation of the CSTS-PS is intended to be used with four to twelve titanium alloy screws which accompany the device.

The CSTS-PS is for single-implant use only. An explanted implant must never be re-implanted. Stresses and fracture, even though not noticeable by visual inspection, may have been created during initial implantation.

Following implantation, the product number and manufacturing lot number of the device that has been implanted must be reported in the patient's surgical file.

#### POTENTIAL ADVERSE EVENTS:

- Potential adverse events may be related to surgery in general, spine surgery specifically or the device. These may include, but are not limited to the following:
  - Adverse events related to any surgery: reactions to anesthesia, the anesthetic or other medications; bleeding; infection; ileus; blood vessel damage; nerve or soft tissue damage; atelectasis; pneumonia; hematoma; seroma; wound dehiscence or incisional hernia; urologic problems; embolism; anemia; colitis; thrombophlebitis; heart attack; stroke; or death.
  - Adverse events related specifically to spine surgery: dorsal tear and CSF leak; nerve damage leading to radiculopathy, myelopathy, paraparesis, paresis or paralysis; meningitis; vertebral body damage or fracture; ligament damage; fractured sacrum; or retrograde ejaculation.
  - Adverse events related to the device: implant crack or fracture, failure to achieve fusion, implant migration, dislodgement, or metal sensitivity to a foreign body, including possible tumor formation. Additional surgery may be necessary for implant removal, repositioning or replacement. Additional stabilization at the implanted level or surgery at another disc level may be necessary if non-union or anatomic change at an adjacent level develops.

#### POSTOPERATIVE CARE:

- The physician's postoperative directions and warnings to the patient and the corresponding patient compliance are extremely important.
- Detailed instructions on the use and limitations of the device must be given to the patient. The patient must be warned that loosening, and / or breakage of the device(s) are complications which may occur as result of early or excessive weight-bearing, muscular activity or sudden jolts or shock to the spine.
- The patient must be advised not to smoke or consume alcohol during period of the bone fusion process.
- The patient must be advised of the inability to bend at the point of spinal fusion and taught to compensate for this permanent physical restriction in body motion.
- It is critical that immobilization of union is established and confirmed by roentgenographic examination. If a non-union develops or if the components loosen, migrate, and / or break, the devices must be revised and / or removed immediately before serious injury occurs.
- Any retrieved devices are not to be used in another surgical procedure.

#### MRI SAFETY INFORMATION:

MRI Safety Information	
A patient with the 4WEB Medical Cervical Plating System may be safely scanned under the following conditions. Failure to follow these conditions may result in injury to the patient.	
Name/Identification of Device	4WEB Medical Cervical Plating System
Nominal Value(s) of Static Magnetic Field [T]	1.5 T or 3 T
Maximum Spatial Field Gradient [T/m and gauss/cm]	20 T/m (2000 gauss/cm)
RF Excitation	Circularly Polarized (CP)
RF Transmit Coil Type	Whole body transmit coil, Head RF transmit-receive coil
Maximum Whole Body SAR [W/kg]	2.0 W/kg (Normal Operating Mode)
Limits on Scan Duration	2.0 W/kg whole body average SAR for 6 minutes of continuous RF (a sequence or back to back series/scan without breaks) with a 5 minute resting period between scans for up to an hour of scanning.
MR Image Artifact	The presence of this implant may produce an image artifact of 28mm.
If information about a specific parameter is not included, there are no conditions associated with that parameter.	

#### PACKAGING:

CSTS-PS implants may be provided sterile and are clearly labeled as such in an unopened sterile package provided by 4WEB. The contents are considered sterile unless the package is damaged, opened, or the expiration date on the device label has passed. The integrity of the packaging should be checked to ensure that the sterility of the contents is not compromised.

#### Implants supplied sterilized from 4WEB must not be re-sterilized.

CSTS-PS ancillary instrumentation and implants may be provided non-sterile and are supplied in a tray or caddy that is used for steam sterilization prior to use in the operating room. In the case of instrument and implant restock, individual items will be sent in a plastic bag labeled for that device.

Storage conditions must maintain the integrity of the implants, associated ancillary instruments and their respective packaging. The condition of all implants and instruments must be checked before use. Damaged products must not be used and should be returned to 4WEB.

#### INFORMATION FOR CLEANING AND STERILIZATION OF SURGICAL INSTRUMENTS AND IMPLANTS:

CAUTION: THESE INSTRUCTIONS DO NOT APPLY TO SINGLE-USE DEVICES.

The instruments used to implant the CSTS-PS do not have an indefinite functional life. All reusable instruments are subjected to repeated stresses related to bone contact, impaction, routing, cleaning, and sterilization processes. Instruments should be carefully inspected before use to ensure that they are fully functional.

All ancillary instruments of the CSTS-PS and certain implants are delivered non-sterile and therefore, must be decontaminated, cleaned and sterilized prior to surgical use. Decontamination reduces the population of microorganisms and facilitates the subsequent cleaning stage. Strict compliance with the instructions for use pertaining to decontamination and cleaning is mandatory, particularly the concentration and exposure time requirements. Thorough rinsing with water must be conducted following decontamination and cleaning.

- Decontamination: Each hospital must use their own validated decontamination procedures.
- Cleaning: Wash instruments in a LANCER type (or equivalent) washing machine with the appropriate cleaning products, rise, and dry. Any product which may alter the material is prohibited, i.e. bleach, formalin, hypochlorite solutions, saline solution, etc.

The reprocessing instructions provided have been validated as being capable of preparing reusable 4WEB instruments. It is the responsibility of the reprocessor to ensure that the reprocessing is actually performed using appropriate equipment, materials, and personnel to achieve the desired result. This normally requires validation and routine monitoring of the process. Any deviation by the reprocessor from these instructions should be evaluated for effectiveness and potential adverse consequences.

#### WARNINGS:

- Follow the instructions and warnings issued by the suppliers of any cleaning and disinfection agents and equipment used.
- Do not exceed 140° C (284° F) during reprocessing steps.
- Highly alkaline conditions can damage products with aluminum parts.
- Avoid exposure to hypochlorite solutions, as these will promote corrosion.

- For instruments produced by another manufacturer, reference the manufacturer's instructions for use.
- Care should be taken to remove any debris, tissue or bone fragments that may collect on the instrument.

#### LIMITATIONS ON REPROCESSING:

- End of useful life is generally determined by wear or damage in surgical use.
- Carefully inspect instruments between uses to verify proper functioning and configuration.
- Damaged instruments must be replaced to prevent potential patient injury such as loss of metal fragments into the surgical site.

#### DECONTAMINATION CONSIDERATIONS – CREUTZFELDT-JAKOB DISEASE (CJD):

- For patients with, or suspected with, Creutzfeldt-Jakob disease (CJD), variant CJD or other transmissible spongiform encephalopathy (TSE) and related infections, it is recommended to treat the patient using single-use instruments.

#### REPROCESSING INSTRUCTIONS

##### CARE AT THE POINT OF USE:

- Use purified water obtained via ultra-filtration, RO, DI and/or distilled.
- Thoroughly clean instruments as soon as possible after use. If cleaning must be delayed, immerse instruments in a compatible pH neutral detergent solution and purified water to prevent drying and encrustation of surgical soil.
- Avoid prolonged exposure to saline to minimize the chance of corrosion.
- Remove excessive soil with a disposable low lint wipe.

##### CLEANING:

##### ALL INSTRUMENTS

- No instruments provided with the CSTS-PS require disassembly prior to cleaning.
- Prepare an enzymatic cleaning solution in accordance with the manufacturer's instructions (1oz Enzol, or equivalent, per gallon of purified water). **Note:** the enzyme solution must be changed on a regular basis.
- Soak soiled instruments for 5 minutes in the enzymatic solution.
- Use a soft bristle brush to remove all traces of blood and debris; pay close attention to any hard-to-reach areas, textured surfaces, or crevices.
- Rinse the instrument thoroughly with purified water.
- Dry the instrument immediately after final rinse.
- Clean the instruments using a validated washer disinfectant and the following parameters:

Process	Time	Reagent	Water Temperature
Pre-wash	2 min	Not applicable	Cold water
Enzyme wash	2 min	Enzol (2oz/gal) or equivalent	Hot water
Wash 1	2 min	Valsure Neutral (1oz/gal) or equivalent	65.5°C
Rinse 1	2 min	Not applicable	Hot water
Thermal rise	1 min	Not applicable	90°C
Drying*	6 min	Not applicable	98.8°C

\*Use lint-free cloths to dry residual water and filtered pressurized air at < 20 psi

- Place heavier instruments on the bottom of containers. Do not place heavy instruments on top of delicate instruments.
- For instruments with concave surfaces, such as curettes, place instrument with the concave surface facing downward to facilitate draining.

##### CLEANING INSPECTION:

- Inspect all instruments to verify that all visible debris is removed during cleaning and prior to sterilization. If debris is still visible after cleaning, repeat the cleaning process.

##### MAINTENANCE INSPECTION AND FUNCTIONAL TESTING:

- Visually inspect all instruments to ensure no damage and wear.
- Ensure there are no cracked handles and shafts are secure in handles.
- Ensure long instruments are free of any bending and distortion.
- Ensure instrument tips are free of defects or burrs.
- Ensure complex instruments with moving parts function appropriately.

##### WRAPPING:

##### WRAPPING TECHNIQUE

- Use instrument trays to contain instruments that are provided in sets.
- Biological or Chemical Indicators (Bis or Cis) used for monitoring the performance of sterilization processes should be placed in the middle racks within wrapped trays. They should be tested according to the BI or CI manufacturer's directions.
- Double wrap instruments in accordance with local procedures, using standard wrapping techniques such as those described in ANSI/AAMI ST79.
- Use only FDA-cleared wraps.
- Label the contents of the wrapped tray using an indelible marker or other sterilization compatible label system.
- Allow 1 inch of free space between the instrument tray and the inside of the container lid for effective processing.

##### STERILIZATION:

- Use a validated, properly maintained and calibrated steam sterilizer.
- Effective steam sterilization can be achieved using the following cycle to achieve an SAL of 10<sup>-6</sup>:

Cycle	Temperature	Duration	Dry Time	Preconditioning pulses
Prevacuum	132°C (270°F)	4 min	60 min	4
*Prevacuum	134°C (273°F)	3 min	60 min	4

\* The Prevacuum 134° C Sterilization Cycle is not considered by the Food and Drug Administration to be a standard sterilization cycle. It is the end user's responsibility to use only sterilizers and accessories (such as sterilization wraps, sterilization pouches, chemical indicators, biological indicators, and sterilization cassettes) that have been cleared by the Food and Drug Administration for the selected sterilization cycle specifications (time and temperature).

**NOTE: STERILIZATION DOES NOT REPLACE DECONTAMINATION OR CLEANING. ONLY A CLEAN PRODUCT CAN BE CORRECTLY STERILIZED. ONLY STERILE IMPLANTS AND INSTRUMENTS MAY BE USED FOR SURGERY.**

##### PRODUCT COMPLAINTS:

Any healthcare professional (e.g. a surgeon using the product) who has a complaint or who has experienced any dissatisfaction in the quality, identity, reliability, safety, efficacy, and/or performance of any CSTS-PS products should notify 4WEB, or, where applicable, their distributor.

##### SURGICAL TECHNIQUE MANUAL:

To receive additional copies of the Surgical Technique Manual, contact your local sales representative or the company at the address below.

##### USA:

4WEB  
2801 Network Blvd, Suite 620  
Frisco, TX USA 75034  
T +1(800) 285-7090  
F +1(972) 488-1816

##### OUTSIDE USA:

4WEB EU B.V.  
Industrieweg 13b  
1566JN Assendelft, The Netherlands  
T +31 20 708 45 45  
F +31 20 708 45 65

Standard: ISO 15223-1, Medical Devices - Symbols to be used with medical device labels, labelling and information to be supplied.				
Symbol	Ref. Number	Title	Description of Symbol	
	5.4.4	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.	
	5.4.2	Do not re-use	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.	
	5.4.3	Consult instructions for use	Indicates the need for the user to consult the instructions for use.	
	5.1.5	Lot number	Indicates the manufacturer's lot number so that a specific medical device can be identified.	
	5.1.6	Catalog number	Indicates the manufacturer's catalog number so that the medical device can be identified.	
	5.1.1	Manufacturer	Indicates the medical device manufacturer, as defined in EU Directives 90/385/EEC, 93/42 EEC, and 98/79 EC.	
	5.1.3	Date of manufacture	Indicates the date when the medical device was manufactured.	
	5.1.4	Use-by date	Indicates the date after which the medical device is not to be used.	
	5.2.4	Sterile	Sterilized using irradiation. Indicates a medical device that has been sterilized using irradiation.	
	5.2.6	Do not re-sterilize	Indicates a medical device that is not to be re-sterilized.	
	5.2.8	Do not use if package is damaged	Indicates a medical device that should not be used if the package has been damaged or opened.	
	5.2.12	Double Sterile Barrier System	Indicates two sterile barrier systems.	
	5.7.7	Medical Device	Indicates the item is a medical device.	
	5.7.10	Unique Device Identifier	Indicates a carrier that contains unique device identifier information.	
	ASTM F2503-13	MR Conditional	Medical device that has been demonstrated to pose no known hazards in a specified MR environment with specified conditions of use.	
	21 CFR 801.109	Prescription only	Requires prescription in the United States.	